

Evoked EEG Electrode Kit

NE-132B

General

The NE-132B evoked EEG electrode kit is used for evoked potential EEG examination.

Safety Information

- ⚠ WARNING** A warning alerts the user to possible injury or death associated with the use or misuse of the instrument.
- ⚠ CAUTION** A caution alerts the user to possible injury or problems with the instrument associated with its use or misuse such as instrument malfunction, instrument failure, damage to the instrument, or damage to other property.

Pay attention to all safety information in the Operator's Manual or Installation Guide.

⚠ WARNING

When performing an MR examination, remove all electrodes from the patient. Failure to follow this warning may cause skin burns on the patient. For details, refer to the MR equipment manual.

⚠ CAUTION

Do not attach the electrode on a wound or inflamed skin. Redness or irritation may appear.

⚠ CAUTION

Do not use this electrode for electric stimulation.

⚠ CAUTION

Do not clean the electrode with hot water or chloric bleaching agent.

⚠ CAUTION

Failure to follow the instructions below may deteriorate the electrode.

- Do not sterilize the electrode by autoclaving or ethylene oxide gas.
- Do not disinfect the electrode with cresol or glutaraldehyde.

⚠ CAUTION

Do not store this electrode in an atmosphere with sulfur. It will deteriorate the electrode.

This Safety and Performance Information is an extract from the general and safety information sections of the most recent edition of Operator's Manual or Installation Guide. Therefore, the contents of your Operator's Manual or Installation Guide may differ from those of this Safety and Performance Information. For detailed operating procedures, follow the instructions of your Operator's Manual or Installation Guide.

 Manufacturer
NIHON KOHDEN CORPORATION
1-31-4 Nishiochiai, Shinjuku-ku,
Tokyo 161-8560, Japan
Phone +81 3-5996-8041
<https://www.nihonkohden.com/>

EC REP European Representative
NIHON KOHDEN EUROPE GmbH
Raiffeisenstrasse 10, 61191 Rosbach, Germany
Phone +49 6003-827-0 Fax +49 6003-827-599